

인정번호 : KC02-140 (1/1)

대성계전주식회사

☐ 공인유효기간 : 2002. 5. 2 ~ 2007. 5. 1

☐ 주 소 : (626-230)

289-22

☐ 답 당 부 서 : 교정과

☐ 품 질 책임자 : 진치환

☐ 전 화 번 호 : 055-385-7894

☐

: 055-384-0152

☐ 인 정 범 위 : (1 4)

(10)

분류코드	교정항목 ()	범 위	최고측정능력(±) (95 %)	기술 책임자
1040020	기체용면적유량계 (DSM-G-01)	(1) 0.006 m ³ /h ~ 250 m ³ /h	(2) 2.40 × 10 ⁻³	김수환
1040030	기체용용적유량계 (DSM-G-02, G-02-1)	0.006 m ³ /h ~ 250 m ³ /h 8 m ³ /h ~ 3 500 m ³ /h	2.40 × 10 ⁻³ 2.66 × 10 ⁻³	
1040040	기체용질량유량계 (DSM-G-03)	0.006 m ³ /h ~ 250 m ³ /h	2.40 × 10 ⁻³	
1040060	기체용터빈유량계 (DSM-G-04, G-04-1)	0.006 m ³ /h ~ 250 m ³ /h 8 m ³ /h ~ 3 500 m ³ /h	2.40 × 10 ⁻³ 2.66 × 10 ⁻³	

주 1 : “0.006 m³/h~ 30 m³/h” 0.006 m³/h 250 m³/h .

주 2 : BMC (2.4 X 10⁻³ = 0.24 %) ; 지수형태로 나타난 상대불확도 값을 표시한 것임.
끝.

기술 평가사 최 용 문 ()

평 가 반 장 김 태 민 ()

Accreditation No : KC02-140 (1/1)

DAESUNG MEASURING Co., Ltd.

☐ Valid Until : 2002. 5. 2 ~ 2007. 5. 1

☐ Address : (Post No. 626-230) 289-22, YOUSAN-DONG, YANGSAN-SI, KYUNGNAM,
KOREA

☐ Lab. Name : CALIBRATION Dept.

☐ Quality Manager : Jin Chi Hoan

☐ Telephone : 055-385-7894

☐ Fax : 055-384-0152

☐ Accredited Scope : Fluid Flow Field (1 Field 4 Item)

(10 Fluid Flow Field)

Field Code	Item of Calibration (Method of Calibration)	Range	BMC (The Confidence Level is 95 %)	Technical Manager
1040020	Rotameter for Gas (DSM-G-01)	^(*) 0.006 m ³ /h ~ 250 m ³ /h	2.40×10^{-3}	KIM, Soo Hoan
1040030	Gas Volumetric Meter (DSM-G-02, G02-1)	0.006 m ³ /h ~ 250 m ³ /h 8 m ³ /h ~ 2 500 m ³ /h	2.40×10^{-3} 2.66×10^{-3}	
1040040	Mass Flow Meter for Gas (DSM-G-03)	0.006 m ³ /h ~ 250 m ³ /h	2.40×10^{-3}	
1040060	Gas Turbine Meter (DSM-G-04, G04-1)	0.006 m ³ /h ~ 250 m ³ /h 8 m ³ /h ~ 2 500 m ³ /h	2.40×10^{-3} 2.66×10^{-3}	

* : "0.006 m³/h ~ 50 m³/h" means flow rate between 0.006 m³/h and 250 m³/h.

End.